Hemochron Signature Elite - High Range (HR) ACT+ Electronic Quality Control (EQC) Record Form

Instructions : EQC to be performed every 8 hours when instrument is in use. All parameters must pass (criteria listed below table) prior to use for patient testing. Document any issues with instrument, cuvette, EQC etc. noted during operation. Include remdial actions taken when QC does not pass.

Date	Time	Operator ID	Cuvette lot In use this day	Instrument #1: SN11734				Instrument #2: SN11749				All QC Pass?	Comments
				Battery	Temp	Low	High	Battery	Temp	Low	High		

EQC Acceptance Criteria:

Battery	8 hour min charge required
Temp	37 °C ±1.0 °C
Low EQC	**29 - 31**
High EQC	**499 - 501**