

Urine Dipstick: Form A1 - Quality Control Current Lot Information

Urisys 1100/ Manual Dip Test Strip Results – Quality Control

Chemstrip 10A (Roche) Urine Test Strip: Lot # _____ Exp: Date _____

Effective: Nov 05.2016

In use: Nov 05.2016

Refrigerated: One month open vial expiry.Room Temperature: 10 day expiry.

Please date vials.

Quality Control BIO-RAD qUANTify Control

NORMAL LEVEL 1
Bio - Rad

Lot#: 78751 Exp:2017.10.14

ABNORMAL LEVEL 2
Bio - Rad

Lot#:78752 Exp:2017.10.14

Refractometer Atago
Specific Gravity

1.010 - 1.025

1.005 – 1.015

1.0175 +/- 0.007

1.010 +/- 0.002

Constituent	Units	Chemstrip Visual	URISYS 1100	Chemstrip Visual	URISYS 1100
Specific Gravity	Test Strip	1.015-1.030	1.015-1.030	1.000-1.015	1.005-1.020
pH		5-6	5-7	6-8	6.5-8
Leukocytes	Leuco/uL	Negative	Negative	Tr - 2+ 25 - 100	2 - 3 + 100-500
Nitrite		Negative	Negative	Positive	Positive
Protein	g/L	Negative	Negative	2-4 + 0.3-5.0 g/L	2-4+ 0.25-5.0 g/L
Glucose	mmol/L	Normal	Normal	2 - 4 + 6 - >56	2 - 4 + 6->56
Ketones	mmol/L	Negative	Negative	1 - 3 + 0.5-5	1 - 4+ 0.5-15
Urobilinogen	mmol/L	Normal	Normal	Normal – 200 umol/L	Normal – 140 umol/L
Bilirubin	umol/L	Negative	Negative	2-3 + 50-100	2-3+ 50-100
Blood	Ery/uL	Negative	Negative	3-4 +	3-5+

Medical Approval: Dr Benjamin Jung

Version: 1.3

Folder Name: CW\Point of Care\Urine Dipstick

This is a controlled document for CW use only. Any printed copies are uncontrolled unless specified. Please refer to Lab QMS

Medical Approval Date: 28 Dec 2016

Implementation Date: 4/30/2017 11:08:53 PM

				50-250	50-250
--	--	--	--	--------	--------

Medical Approval: Dr Benjamin Jung

Version: 1.3

Folder Name: CW\Point of Care\Urine Dipstick

Medical Approval Date: 28 Dec 2016

Implementation Date: 4/30/2017 11:08:53 PM

This is a controlled document for CW use only. Any printed copies are uncontrolled unless specified. Please refer to Lab QMS

REVISION & APPROVAL LOG

Version	Revision Type	Description of Change	Revision Date	Technical Approval	Medical Approval
1.0		New document	Nov 2008	Elvira Kozak	Dr. Cathy Halstead
1.1	Minor	Revision	05 Nov 2016	Elvira Kozak	
1.2	Minor	Document title and number change. Upload to QMS document control	28 Dec 2016		Dr. Benjamin Jung
1.3	Minor	Corrected Room Temp QC expiry	Apr 30, 2017	Calvin Lee	

Attention: This document is published on the BCCW SharePoint website

Revisions made to this document require an update to the corresponding document published on BCCW SharePoint website

Medical Approval: Dr Benjamin Jung

Version: 1.3

Folder Name: CW\Point of Care\Urine Dipstick

This is a controlled document for CW use only. Any printed copies are uncontrolled unless specified. Please refer to Lab QMS

Medical Approval Date: 28 Dec 2016

Implementation Date: 4/30/2017 11:08:53 PM