

Urine Dipstick Emergency Dept: Form C1 - **Abnormal Quality Control** Record Sheet

DAILY Urine Quality Control (UQC) –BioRad qUAntify ABNORMAL LEVEL 2
 Note: Store refrigerated - open Vial Expiry is 1 month. Room Temperature stability is 10 days. Obtain from Chemistry Lab walk-in fridge – C&W
 Point of Care Testing 604-875-2345 ext: 7521

Chemstrip 10A Lot # _____							UQC Lot # _____ Expiration date: _____					
Date	Time	Specific Gravity	pH	LEU	NIT	PRO g/L	GLU mmol/L	KET mmol/L	URO mmol/L	BILI umol/L	BLOOD	RN
Read @	XX sec	60	60	60-120	60	60	60	60	60	60	60	Operator ID
AbNormal Level		1.005-1.015	6 -8	Tr -2 + 25 - 100	POS	2-4 + 0.3-5.0	2-4 + 6 - >56	1-3 + 0.5-15	1-3 + Norm-200	2-3 + 50-100	3-4 + 50-250	

REVISION & APPROVAL LOG

Version	Revision Type	Description of Change	Revision Date	Technical Approval	Medical Approval
1.0		New document	Nov 2008	Elvira Kozak	Dr. Cathy Halstead
1.1	Minor	Revision	05 Nov 2016	Elvira Kozak	
1.2	Minor	Document title and number change. Upload to QMS document control	28 Dec 2016		Dr. Benjamin Jung
1.3	Minor	Corrected QC stability	Apr 30, 2017	Calvin Lee	

Attention: This document is published on the BCCW SharePoint website

Revisions made to this document require an update to the corresponding document published on BCCW SharePoint website