

Urine Dipstick Medical Mobile Unit: Form B2 - **Normal Quality Control** Record Sheet**DAILY Urine Quality Control (UQC) – BioRad qUAntify NORMAL LEVEL***Note: Store at Room temperature. Open Vial Expiry is 1 months. Obtain from 2J8 – C&W Point of Care Testing 604-875-2345 ext: 7521*

Chemstrip 10A Lot #							UQC Lot # 78751				Expiry : 2017-10-14	
Date	Time	Specific Gravity	pH	LEU	NIT	PRO	GLU	KET	URO	BILI	BLOOD	RN
Read @	XX sec	60	60	60-120	60	60	60	60	60	60	60	Operator ID
Normal Level UQC Results		1.015-1.030	5.0-7.0	NEG	NEG	NEG	Norm	NEG	Norm	NEG	NEG	

REVISION & APPROVAL LOG

Version	Revision Type	Description of Change	Revision Date	Technical Approval	Medical Approval
1.0		New document	Nov 2008	Elvira Kozak	Dr. Cathy Halstead
1.1	Minor	Revision	05 Nov 2016	Elvira Kozak	
1.2	Minor	Document title and number change. Upload to QMS document control	28 Dec 2016		Dr. Benjamin Jung

Attention: This document is published on the BCCW SharePoint website

Revisions made to this document require an update to the corresponding document published on BCCW SharePoint website